

Document No: RN-0145

CE DECLARATION OF CONFORMITY

Hanna Instruments SRL - Nusfalau Science Park Romania, Salaj, Nusfalau, No. 1, Hanna Str. Tel.: +40/260/607700; Fax: +40/260/607739

This certifies that the following designated products

HI 2020

are in conformity with the EMC directive 2004/108/EC and its standards, and Low Voltage Directive 2006/95/EC and its standards:

EN 61326-1: 2013 Electrical Equipment for Measurement, Control and Laboratory use - EMC requirements Part 1: General requirements and its standards:

EN 61000-4-2: 2009, EMC - Electrostatic Discharge Immunity Test

EN 61000-4-3: 2006, EMC - Radiated RF, Electromagnetic Field Immunity Test

EN 61000-4-4: 2012, EMC - Electrical Fast Transient/Burst Immunity Test

Class A

4 kV/8 kV contact/air 3 V/m (80 MHz to 1 GHz) 1 kV (5/50 ns, 5 kHz)

EN 61010-1: 2010 (IEC 61010-1: 2010) Safety Requirements for Electrical Equipment for Measurement, Control and Laboratory Use Part 1: General Requirements.

Date of issue: February, 03, 2014

Name and Function; Marius Talpos elect. Eng.



Document No: RN-0146

CE DECLARATION OF CONFORMITY

Hanna Instruments SRL - Nusfalau Science Park Romania, Salaj, Nusfalau, No. 1, Hanna Str. Tel.: +40/260/607700; Fax: +40/260/607739

This certifies that the following designated products

HI 2030

are in conformity with the EMC directive 2004/108/EC and its standards, and Low Voltage Directive 2006/95/EC and its standards:

EN 61326-1: 2013 Electrical Equipment for Measurement, Control and Laboratory use - EMC requirements Part 1: General requirements

and its standards:

EN 61000-4-2: 2009, EMC - Electrostatic Discharge Immunity Test

EN 61000-4-3: 2006, EMC - Radiated RF, Electromagnetic Field Immunity Test

EN 61000-4-4: 2012, EMC - Electrical Fast Transient/Burst Immunity Test

Class A

4 kV/8 kV contact/air 3 V/m (80 MHz to 1 GHz)

1 kV (5/50 ns, 5 kHz)

EN 61010-1: 2010 (IEC 61010-1: 2010) Safety Requirements for Electrical Equipment for Measurement, Control and Laboratory Use Part 1: General Requirements.

Date of issue:

February, 03, 2014

Name and Function:

Marius Talpos elect. Eng.



Document No: RN-0168

CE DECLARATION OF CONFORMITY

Hanna Instruments SRL - Nusfalau Science Park Romania, Salaj, Nusfalau, No. 1, Hanna Str. Tel.: +40/260/607700; Fax: +40/260/607739

This certifies that the following designated products

HI 2040

are in conformity with the EMC directive 2004/108/EC and its standards, and Low Voltage Directive 2006/95/EC and its standards:

EN 61326-1: 2013 Electrical Equipment for Measurement, Control and Laboratory use - EMC requirements Part 1: General requirements and its standards:

EN 61000-4-2: 2009, EMC - Electrostatic Discharge Immunity Test

EN 61000-4-3: 2006, EMC - Radiated RF, Electromagnetic Field Immunity Test

EN 61000-4-4: 2012, EMC - Electrical Fast Transient/Burst Immunity Test

EN 61010-1: 2010 (IEC 61010-1: 2010) Safety Requirements for Electrical Equipment for Measurement, Control and Laboratory Use Part 1: General Requirements. Class A

4 kV/8 kV contact/air 10 V/m (80 MHz to 1 GHz) 1 kV (5/50 ns, 5 kHz)

Date of issue: June, 30, 2014

Name and Function: Marius Talpos Elect.Eng

Rev. 0.1 2014-06-30